

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061419\
 Data File : VV011524.D
 Acq On : 15 Jun 2019 08:01
 Operator : SY/MD
 Sample : VV0614WBL02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK32

Quant Time: Jun 17 04:56:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 17 04:30:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	169484	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	150286	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56682	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	59389	4.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.80%
7) Chloroethane-d5	1.58	69	45907	5.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.20%
11) 1,1-Dichloroethene-d2	2.12	63	85495	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	3.96	46	186212	51.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.44%
24) Chloroform-d	4.40	84	115710	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
26) 1,2-Dichloroethane-d4	5.08	65	71292	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.09	84	226918	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.11	67	72780	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.36	98	183348	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	7.66	79	21011	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.14	63	145750	53.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.66%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	51003	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	58189	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061419\
Data File : VV011524.D
Acq On : 15 Jun 2019 08:01
Operator : SY/MD
Sample : VV0614WBL02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK32

Quant Time: Jun 17 04:56:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 17 04:30:15 2019
Response via : Initial Calibration

