

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV061419\
 Data File : VV011503.D
 Acq On : 14 Jun 2019 19:13
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV36

Quant Time: Jun 15 01:22:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 18:58:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.446	0.436	2.2	99	0.00
3 T	Chloromethane	0.483	0.480	0.6	105	0.00
4 S	Vinyl Chloride-d3	0.377	0.366	2.9	101	0.00
5 T	Vinyl chloride	0.477	0.448	6.1	96	0.00
6 T	Bromomethane	0.200	0.206	-3.0	112	0.00
7 S	Chloroethane-d5	0.239	0.234	2.1	105	0.00
8 T	Chloroethane	0.231	0.220	4.8	107	0.00
9 T	Trichlorofluoromethane	0.526	0.528	-0.4	100	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.300	0.295	1.7	100	0.00
11 S	1,1-Dichloroethene-d2	0.719	0.690	4.0	100	0.00
12 T	1,1-Dichloroethene	0.311	0.290	6.8	96	0.00
13 T	Acetone	0.072	0.064	11.1	95	0.00
14 T	Carbon disulfide	0.841	0.797	5.2	98	0.00
15 T	Methyl Acetate	0.190	0.168	11.6	91	0.00
16 T	Methylene chloride	0.353	0.332	5.9	99	0.00
17 T	Methyl tert-butyl Ether	0.932	0.891	4.4	96	0.00
18 T	trans-1,2-Dichloroethene	0.352	0.343	2.6	98	0.00
19 T	1,1-Dichloroethane	0.727	0.700	3.7	97	0.00
20 S	2-Butanone-d5	0.106	0.108	-1.9	103	0.00
21 T	2-Butanone	0.113	0.110	2.7	98	0.00
22 T	cis-1,2-Dichloroethene	0.398	0.384	3.5	100	0.00
23 T	Bromochloromethane	0.157	0.151	3.8	100	0.00
24 S	Chloroform-d	0.726	0.713	1.8	102	0.00
25 T	Chloroform	0.730	0.698	4.4	96	0.00
26 S	1,2-Dichloroethane-d4	0.401	0.405	-1.0	103	0.00
27 T	1,2-Dichloroethane	0.514	0.487	5.3	97	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
29 T	1,1,1-Trichloroethane	0.622	0.598	3.9	95	0.00
30 T	Cyclohexane	0.643	0.610	5.1	99	0.00
31 T	Carbon tetrachloride	0.503	0.487	3.2	95	0.00
32 S	Benzene-d6	1.492	1.451	2.7	101	0.00
33 T	Benzene	1.633	1.589	2.7	97	0.00
34 T	Trichloroethene	0.416	0.396	4.8	95	0.00
35 T	Methylcyclohexane	0.594	0.593	0.2	101	0.00
36 S	1,2-Dichloropropane-d6	0.474	0.459	3.2	101	0.00
37 T	1,2-Dichloropropane	0.435	0.395	9.2	95	0.00
38 T	Bromodichloromethane	0.494	0.471	4.7	95	0.00
39 T	cis-1,3-Dichloropropene	0.547	0.519	5.1	94	0.00
40 T	4-Methyl-2-pentanone	0.310	0.298	3.9	96	0.00
41 S	Toluene-d8	1.330	1.344	-1.1	105	0.00
42 T	Toluene	1.692	1.671	1.2	98	0.00
43 S	trans-1,3-Dichloropropene-d	0.158	0.160	-1.3	107	0.00
44 T	trans-1,3-Dichloropropene	0.419	0.422	-0.7	101	0.00
45 T	1,1,2-Trichloroethane	0.290	0.274	5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV061419\
 Data File : VV011503.D
 Acq On : 14 Jun 2019 19:13
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV36

Quant Time: Jun 15 01:22:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 18:58:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.091	0.092	-1.1	102	0.00
47 T	Tetrachloroethene	0.297	0.285	4.0	99	0.00
48 T	2-Hexanone	0.218	0.215	1.4	97	0.00
49 T	Dibromochloromethane	0.292	0.280	4.1	95	0.00
50 T	1,2-Dibromoethane	0.262	0.261	0.4	99	0.00
51 T	Chlorobenzene	1.039	1.004	3.4	98	0.00
52 T	Ethylbenzene	1.778	1.754	1.3	99	0.00
53 T	m,p-xylene	0.650	0.652	-0.3	98	0.00
54 T	o-xylene	0.653	0.659	-0.9	99	0.00
55 T	Styrene	1.084	1.088	-0.4	99	0.00
56 T	Isopropylbenzene	1.663	1.664	-0.1	100	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.355	0.345	2.8	102	0.00
58 T	1,1,2,2-Tetrachloroethane	0.355	0.338	4.8	97	0.00
59	1,2,3-Trichloropropane	0.271	0.258	4.8	97	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
61 T	Bromoform	0.319	0.301	5.6	98	0.00
62 T	1,3-Dichlorobenzene	1.621	1.579	2.6	99	0.00
63 T	1,4-Dichlorobenzene	1.652	1.590	3.8	97	0.00
64 S	1,2-Dichlorobenzene-d4	0.998	1.006	-0.8	105	0.00
65 T	1,2-Dichlorobenzene	1.632	1.582	3.1	98	0.00
66 T	1,2-Dibromo-3-chloropropane	0.119	0.112	5.9	98	0.00
67	1,3,5-Trichlorobenzene	1.184	1.126	4.9	96	0.00
68 T	1,2,4-trichlorobenzene	0.951	0.948	0.3	101	0.00
69	Naphthalene	1.870	1.821	2.6	100	0.00
70 T	1,2,3-Trichlorobenzene	0.922	0.911	1.2	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0