

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061422\
 Data File : VW026229.D
 Acq On : 14 Jun 2022 12:09
 Operator : SY/MD
 Sample : N3273-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B90

Quant Time: Jun 15 05:48:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 05:46:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	149170	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	143052	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55793	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	80141	4.108	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.200%
7) Chloroethane-d5	1.574	69	68688	4.470	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	2.111	63	101035	3.221	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.400%
20) 2-Butanone-d5	3.909	46	85155	58.512	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.020%
24) Chloroform-d	4.355	84	129713	4.689	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	5.037	65	52671	4.722	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.053	84	238311	4.273	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
36) 1,2-Dichloropropane-d6	6.072	67	75098	4.670	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.317	98	179919	3.623	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	7.625	79	17799	3.667	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.400%
46) 2-Hexanone-d5	8.092	63	79595	52.664	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.320%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42288	4.413	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	60024	4.799	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.317	62	1930	0.124	ug/L #	42
9) Trichlorofluoromethane	1.757	101	392742	18.709	ug/L	98
12) 1,1-Dichloroethene	2.121	96	2647	0.232	ug/L #	1
13) Acetone	2.192	43	45344	54.570	ug/L	95
19) 1,1-Dichloroethane	3.201	63	2347	0.112	ug/L	96
22) cis-1,2-Dichloroethene	3.918	96	56516	5.137	ug/L	97
25) Chloroform	4.381	83	23393	1.066	ug/L	99
31) Carbon tetrachloride	4.835	117	1650	0.109	ug/L	85
33) Benzene	5.114	78	3284	0.076	ug/L	100
34) Trichloroethene	5.915	95	108086	9.250	ug/L	99
45) 1,1,2-Trichloroethane	7.844	97	6300	0.893	ug/L	90
47) Tetrachloroethene	7.976	164	1218396	145.202	ug/L	98
51) Chlorobenzene	8.886	112	13106	0.445	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061422\
Data File : VV026229.D
Acq On : 14 Jun 2022 12:09
Operator : SY/MD
Sample : N3273-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B90

Quant Time: Jun 15 05:48:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 05:46:56 2022
Response via : Initial Calibration

