

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061422\
 Data File : VW026242.D
 Acq On : 14 Jun 2022 18:06
 Operator : SY/MD
 Sample : N3272-11DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B55DL

Quant Time: Jun 15 05:52:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 05:46:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	176043	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	167283	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68738	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.314	65	94693	4.113	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.200%
7) Chloroethane-d5	1.574	69	79491	4.384	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.111	63	131430	3.551	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	3.918	46	131055	76.305	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	152.600%#
24) Chloroform-d	4.355	84	159916	4.898	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.040	65	71263	5.414	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.200%
32) Benzene-d6	5.053	84	316051	4.846	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	6.072	67	98202	5.222	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.400%
41) Toluene-d8	7.317	98	253711	4.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.625	79	27396	4.827	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.091	63	99339	56.207	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.420%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	56389	5.033	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	79513	5.160	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.200%
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.760	101	7027	0.284	ug/L	96
16) Methylene chloride	2.513	84	14618	1.054	ug/L	93
22) cis-1,2-Dichloroethene	3.924	96	1483	0.114	ug/L	93
34) Trichloroethene	5.915	95	111530	8.162	ug/L	98
47) Tetrachloroethene	7.979	164	4425	0.451	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061422\
Data File : VV026242.D
Acq On : 14 Jun 2022 18:06
Operator : SY/MD
Sample : N3272-11DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B55DL

Quant Time: Jun 15 05:52:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 05:46:56 2022
Response via : Initial Calibration

