

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061422\
 Data File : VW026245.D
 Acq On : 14 Jun 2022 19:17
 Operator : SY/MD
 Sample : N3272-15DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B59DL

Quant Time: Jun 15 05:52:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 05:46:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	168887	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	165721	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68927	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.314	65	94718	4.289	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.800%	
7) Chloroethane-d5	1.574	69	82877	4.764	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.200%	
11) 1,1-Dichloroethene-d2	2.114	63	127473	3.590	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.800%	
20) 2-Butanone-d5	3.912	46	106415	64.584	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	129.160%	
24) Chloroform-d	4.359	84	160347	5.119	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.400%	
26) 1,2-Dichloroethane-d4	5.040	65	69515	5.505	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.000%	
32) Benzene-d6	5.053	84	312919	4.843	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	6.072	67	96853	5.199	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.000%	
41) Toluene-d8	7.317	98	247423	4.301	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.000%	
43) trans-1,3-Dichloroprop...	7.625	79	26843	4.774	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.400%	
46) 2-Hexanone-d5	8.092	63	99633	56.905	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	113.800%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	57852	5.212	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	82484	5.339	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.800%	
Target Compounds						
9) Trichlorofluoromethane	1.758	101	15149	0.637	ug/L	92
16) Methylene chloride	2.516	84	11711	0.880	ug/L	95
31) Carbon tetrachloride	4.831	117	679	0.039	ug/L #	45
34) Trichloroethene	5.918	95	80013	5.911	ug/L	99
47) Tetrachloroethene	7.976	164	23086	2.375	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061422\
Data File : VV026245.D
Acq On : 14 Jun 2022 19:17
Operator : SY/MD
Sample : N3272-15DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B59DL

Quant Time: Jun 15 05:52:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 05:46:56 2022
Response via : Initial Calibration

