

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061422\
 Data File : VW026259.D
 Acq On : 15 Jun 2022 01:24
 Operator : SY/MD
 Sample : N3275-11 200X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BE3

Quant Time: Jun 15 09:41:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 15 07:11:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	155669	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	153876	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61838	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.314	65	83233	4.089	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.800%
7) Chloroethane-d5	1.574	69	74553	4.650	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	2.114	63	115661	3.533	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.600%
20) 2-Butanone-d5	3.918	46	125460	82.608	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	165.220%#
24) Chloroform-d	4.355	84	151850	5.260	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.200%
26) 1,2-Dichloroethane-d4	5.037	65	69527	5.973	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.400%
32) Benzene-d6	5.053	84	299025	4.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	6.072	67	93045	5.379	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.600%
41) Toluene-d8	7.317	98	230820	4.321	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
43) trans-1,3-Dichloroprop...	7.625	79	26231	5.024	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.400%
46) 2-Hexanone-d5	8.092	63	92520	56.910	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.820%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	55120	5.348	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	78356	5.653	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.000%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.516	84	3283	0.268	ug/L	87
25) Chloroform	4.381	83	6160	0.269	ug/L	100
47) Tetrachloroethene	7.982	164	673	0.075	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061422\
Data File : VV026259.D
Acq On : 15 Jun 2022 01:24
Operator : SY/MD
Sample : N3275-11 200X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE3

Quant Time: Jun 15 09:41:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 15 07:11:20 2022
Response via : Initial Calibration

