

Data File : VV006272.D
Acq On : 15 Jun 2018 10:21
Operator : SY/MD
Sample : VV0615WBL01
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Jun 16 00:02:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 15 23:47:49 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	349338	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	307198	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	139481	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	88630	4.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.00%
7) Chloroethane-d5	1.58	69	73675	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.40%
11) 1,1-Dichloroethene-d2	2.13	63	128540	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	3.98	46	185496	45.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.58%
24) Chloroform-d	4.40	84	151678	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
26) 1,2-Dichloroethane-d4	5.09	65	78115	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
32) Benzene-d6	5.10	84	330875	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
36) 1,2-Dichloropropane-d6	6.12	67	97668	4.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.00%
41) Toluene-d8	7.36	98	314498	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.67	79	37023	3.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.60%
46) 2-Hexanone-d5	8.14	63	153146	35.72	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.44%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	60931	3.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	110118	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds					Qvalue
16) Methylene chloride	2.54	84	9068	0.410 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV006272.D
Acq On : 15 Jun 2018 10:21
Operator : SY/MD
Sample : VV0615WBL01
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Jun 16 00:02:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 15 23:47:49 2018
Response via : Initial Calibration

