

Data File : VV006277.D
Acq On : 15 Jun 2018 12:53
Operator : SY/MD
Sample : J3507-18MSD
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWD0MSD

Quant Time: Jun 16 00:15:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 15 23:47:49 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	300046	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	271734	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	122376	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55104	3.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.20%
7) Chloroethane-d5	1.58	69	56282	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.80%
11) 1,1-Dichloroethene-d2	2.13	63	117600	3.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	3.95	46	217171	62.41	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.82%
24) Chloroform-d	4.41	84	132631	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.09	65	70571	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
32) Benzene-d6	5.11	84	261013	4.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.00%
36) 1,2-Dichloropropane-d6	6.12	67	84872	4.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.40%
41) Toluene-d8	7.37	98	224658	3.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.20%
43) trans-1,3-Dichloropropene-	7.67	79	29592	3.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.40%
46) 2-Hexanone-d5	8.14	63	181866	47.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.90%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	65643	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	96458	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds

					Qvalue
12) 1,1-Dichloroethene	2.14	96	72092	4.187 ug/L	94
13) Acetone	2.20	43	9398	4.984 ug/L	92
33) Benzene	5.15	78	321133	4.286 ug/L	100
34) Trichloroethene	5.96	95	81110	3.966 ug/L	98
42) Toluene	7.44	91	408331	4.938 ug/L	99
51) Chlorobenzene	8.93	112	244090	4.597 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV006277.D
Acq On : 15 Jun 2018 12:53
Operator : SY/MD
Sample : J3507-18MSD
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWD0MSD

Quant Time: Jun 16 00:15:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 15 23:47:49 2018
Response via : Initial Calibration

