

Data File : VV006295.D
Acq On : 15 Jun 2018 22:14
Operator : SY/MD
Sample : J3550-12
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWG3

Quant Time: Jun 16 01:05:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 16 00:57:31 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	279572	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	249423	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	113426	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80046	5.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.80%
7) Chloroethane-d5	1.58	69	69772	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.20%
11) 1,1-Dichloroethene-d2	2.13	63	112870	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	3.96	46	189946	58.59	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.18%
24) Chloroform-d	4.41	84	149023	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.09	65	75181	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.11	84	314562	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.12	67	95603	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.20%
41) Toluene-d8	7.37	98	285326	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.67	79	31612	3.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.40%
46) 2-Hexanone-d5	8.14	63	157551	45.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.50%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	62639	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	107050	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds					Qvalue
13) Acetone	2.21	43	11350	6.460 ug/L	93
21) 2-Butanone	4.05	43	43526	13.618 ug/L	92
34) Trichloroethene	5.96	95	50195	2.674 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV006295.D
Acq On : 15 Jun 2018 22:14
Operator : SY/MD
Sample : J3550-12
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWG3

Quant Time: Jun 16 01:05:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 16 00:57:31 2018
Response via : Initial Calibration

