

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061523\
 Data File : VW031547.D
 Acq On : 15 Jun 2023 22:42
 Operator : SY/MD
 Sample : 03249-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYW1

Quant Time: Jun 16 06:17:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 16 05:51:14 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/16/2023
 Supervised By : Mahesh Dadoda 06/16/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	251861	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	243474	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	119135	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	60451	36.064	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	72.120%		
7) Chloroethane-d5	1.539	69	56897	38.902	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	77.800%		
11) 1,1-Dichloroethene-d2	2.066	65	32405	37.510	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.020%		
21) 2-Butanone-d5	3.799	46	117865	86.839	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	86.840%		
24) Chloroform-d	4.259	84	163076	38.847	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	77.700%		
26) 1,2-Dichloroethane-d4	4.950	65	110964	42.248	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.500%		
32) Benzene-d6	4.970	84	297053	43.345	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.680%		
36) 1,2-Dichloropropane-d6	5.995	67	107581	44.290	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.580%		
41) Toluene-d8	7.249	98	251642	40.013	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	80.020%		
43) trans-1,3-Dichloroprop...	7.558	79	48516	40.972	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.940%		
47) 2-Hexanone-d5	8.030	63	83506	87.071	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.070%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	148120	40.400	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	80.800%		
66) 1,2-Dichlorobenzene-d4	11.567	152	108017	47.565	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.140%		
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.719	101	7694	1.851	ug/L	100
12) 1,1-Dichloroethene	2.076	96	222840m	119.576	ug/L	
17) trans-1,2-Dichloroethene	2.706	96	6448	3.315	ug/L	93
19) 1,1-Dichloroethane	3.121	63	470064	111.961	ug/L	99
20) cis-1,2-Dichloroethene	3.825	96	269149	128.029	ug/L	99
25) Chloroform	4.291	83	29854	6.840	ug/L	99
30) 1,1,1-Trichloroethane	4.523	97	1234676	316.291	ug/L	99
34) Trichloroethene	5.838	95	253066	116.830	ug/L	98
45) 1,1,2-Trichloroethane	7.780	97	5665	2.519	ug/L	93
46) Tetrachloroethene	7.912	164	100978	59.396	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061523\
Data File : VV031547.D
Acq On : 15 Jun 2023 22:42
Operator : SY/MD
Sample : 03249-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXYW1

Quant Time: Jun 16 06:17:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 16 05:51:14 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/16/2023
Supervised By : Mahesh Dadoda 06/16/2023

