

Data File : VV016950.D
Acq On : 16 Jun 2020 10:38
Operator : SY/MD
Sample : L2914-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 17 01:07:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 17 00:55:26 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	328857	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	317236	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	156520	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92024	39.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.32%
7) Chloroethane-d5	1.58	69	70774	50.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.66%
11) 1,1-Dichloroethene-d2	2.12	63	140397	36.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.06%
21) 2-Butanone-d5	3.90	46	168190	97.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.93%
24) Chloroform-d	4.38	84	224357	47.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.10%
26) 1,2-Dichloroethane-d4	5.05	65	164189	51.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.32%
32) Benzene-d6	5.07	84	438387	47.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.88%
36) 1,2-Dichloropropane-d6	6.09	67	138046	48.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.86%
41) Toluene-d8	7.34	98	407086	48.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.02%
43) trans-1,3-Dichloropropene-	7.64	79	67865	45.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.42%
47) 2-Hexanone-d5	8.11	63	95002	76.44	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.44%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	165214	43.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.56%
64) 1,2-Dichlorobenzene-d4	11.65	152	171873	51.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.90%

Target Compounds					Qvalue
25) Chloroform	4.40	83	4318	1.003 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016950.D
Acq On : 16 Jun 2020 10:38
Operator : SY/MD
Sample : L2914-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 17 01:07:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 17 00:55:26 2020
Response via : Initial Calibration

