

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061623\
 Data File : VW031587.D
 Acq On : 16 Jun 2023 19:46
 Operator : SY/MD
 Sample : 03249-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYS3

Quant Time: Jun 17 02:32:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 17 02:08:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	198608	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	202291	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	96722	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	37200	28.143	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	56.280%#		
7) Chloroethane-d5	1.539	69	39681	34.406	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	68.820%#		
11) 1,1-Dichloroethene-d2	2.066	65	20647	30.308	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	60.620%		
21) 2-Butanone-d5	3.799	46	101438	94.775	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	94.780%		
24) Chloroform-d	4.259	84	130533	39.432	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	78.860%		
26) 1,2-Dichloroethane-d4	4.950	65	95054	45.894	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.780%		
32) Benzene-d6	4.969	84	226704	39.814	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	79.620%		
36) 1,2-Dichloropropane-d6	5.995	67	86593	42.907	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	85.820%		
41) Toluene-d8	7.249	98	187463	35.877	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	71.760%#		
43) trans-1,3-Dichloroprop...	7.558	79	40278	40.940	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.880%		
47) 2-Hexanone-d5	8.034	63	70039	87.897	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.900%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	139033	45.641	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.280%		
66) 1,2-Dichlorobenzene-d4	11.567	152	89012	48.279	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.560%		
Target Compounds						Qvalue
5) Vinyl chloride	1.288	62	174380	88.492	ug/L	98
8) Chloroethane	1.555	64	158936	124.877	ug/L	96
12) 1,1-Dichloroethene	2.076	96	6546	4.454	ug/L #	1
17) trans-1,2-Dichloroethene	2.706	96	13521	8.814	ug/L	91
19) 1,1-Dichloroethane	3.121	63	1051134	317.492	ug/L	99
20) cis-1,2-Dichloroethene	3.822	96	486670	293.573	ug/L	97
25) Chloroform	4.285	83	11587	3.367	ug/L	100
27) 1,2-Dichloroethane	5.053	62	20538	7.732	ug/L	96
30) 1,1,1-Trichloroethane	4.522	97	261323	80.573	ug/L	99
33) Benzene	5.030	78	8573	1.273	ug/L	100
34) Trichloroethene	5.844	95	16037	8.911	ug/L	89
46) Tetrachloroethene	7.915	164	6646	4.705	ug/L	95
63) 1,2,4-Trimethylbenzene	10.860	105	3961	0.834	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061623\
Data File : VV031587.D
Acq On : 16 Jun 2023 19:46
Operator : SY/MD
Sample : 03249-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXYS3

Quant Time: Jun 17 02:32:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 17 02:08:04 2023
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 88.492 ug/L
 RT: 1.288 min Scan# 71
 Delta R.T. -0.000 min
 Lab File: VV031587.D
 Acq: 16 Jun 2023 19:46




Instrument :
 MSVOA_V
 ClientSampleId :
 EXYS3


Tgt Ion: 62 Resp: 174380
 Ion Ratio Lower Upper
 62 100
 64 32.3 23.4 43.4



#8
 Chloroethane
 Concen: 124.877 ug/L
 RT: 1.555 min Scan# 160
 Delta R.T. -0.000 min
 Lab File: VV031587.D
 Acq: 16 Jun 2023 19:46

Tgt Ion: 64 Resp: 158936
 Ion Ratio Lower Upper
 64 100
 66 32.1 21.0 39.0





#12

1,1-Dichloroethene

Concen: 4.454 ug/L

RT: 2.076 min Scan# 31

Delta R.T. -0.000 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Instrument :

MSVOA_V

ClientSampleId :

EXYS3





Tgt Ion: 96 Resp: 6546

Ion Ratio Lower Upper

96 100

61 216.8 128.9 239.5

63 600.0 95.3 177.1#



#17

trans-1,2-Dichloroethene

Concen: 8.814 ug/L

RT: 2.706 min Scan# 518

Delta R.T. 0.003 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46




Tgt Ion: 96 Resp: 13521


Ion Ratio Lower Upper

96 100

61 133.3 102.5 190.3

98 59.3 44.1 81.9





#19

1,1-Dichloroethane

Concen: 317.492 ug/L

RT: 3.121 min Scan# 64

Delta R.T. 0.003 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Instrument :





MSVOA_V

ClientSampleId :

EXYS3

Tgt Ion: 63 Resp: 1051134

Ion	Ratio	Lower	Upper
63	100		
65	32.4	22.5	41.7
83	13.9	10.0	18.6



#20

cis-1,2-Dichloroethene

Concen: 293.573 ug/L

RT: 3.822 min Scan# 865




Delta R.T. -0.003 min


Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Tgt Ion: 96 Resp: 486670

Ion	Ratio	Lower	Upper
96	100		
61	135.7	98.6	183.0
98	64.5	45.7	84.9





#25

Chloroform

Concen: 3.367 ug/L

RT: 4.285 min Scan# 1009

Delta R.T. -0.000 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Instrument :

MSVOA_V

ClientSampleId :





EXYS3

Tgt Ion: 83 Resp: 11587

Ion Ratio Lower Upper

83 100

85 64.8 45.3 84.1



#27

1,2-Dichloroethane

Concen: 7.732 ug/L

RT: 5.053 min Scan# 1248

Delta R.T. 0.006 min

Lab File: VV031587.D




Acq: 16 Jun 2023 19:46


Tgt Ion: 62 Resp: 20538

Ion Ratio Lower Upper

62 100

98 9.6 6.6 9.8









#30
1,1,1-Trichloroethane
Concen: 80.573 ug/L
RT: 4.522 min Scan# 1082
Delta R.T. 0.003 min
Lab File: VV031587.D
Acq: 16 Jun 2023 19:46




Instrument :
MSVOA_V
ClientSampleId :
EXYS3


Tgt Ion: 97 Resp: 261323
Ion Ratio Lower Upper
97 100
99 65.0 51.8 77.6
61 40.5 33.0 49.6



#33
Benzene
Concen: 1.273 ug/L
RT: 5.030 min Scan# 1241
Delta R.T. 0.013 min
Lab File: VV031587.D
Acq: 16 Jun 2023 19:46

Tgt Ion: 78 Resp: 8573





#34

Trichloroethene

Concen: 8.911 ug/L

RT: 5.844 min Scan# 1492

Delta R.T. 0.006 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Instrument :

MSVOA_V

ClientSampleId :

EXYS3

Tgt Ion: 95 Resp: 16037





Ion Ratio Lower Upper

95 100

97 62.5 46.8 86.8

132 84.3 66.4 123.2

130 85.4 69.7 129.4



#46

Tetrachloroethene

Concen: 4.705 ug/L

RT: 7.915 min Scan# 2138

Delta R.T. 0.003 min

Lab File: VV031587.D

Acq: 16 Jun 2023 19:46

Tgt Ion:164 Resp: 6646




Ion Ratio Lower Upper


164 100

129 92.4 68.4 127.0

131 95.4 67.6 125.6


166 121.1 91.3 169.5





#63
 1,2,4-Trimethylbenzene
 Concen: 0.834 ug/L
 RT: 10.860 min Scan# 3054
 Delta R.T. -0.000 min
 Lab File: VV031587.D
 Acq: 16 Jun 2023 19:46

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYS3



Tgt Ion:105 Resp: 3961
 Ion Ratio Lower Upper
 105 100
 120 43.7 33.4 50.2

