

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
 Data File : VV038801.D
 Acq On : 16 Jun 2025 12:07
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV247

Quant Time: Jun 16 12:36:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 12:04:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6656	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6213	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2741	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.287	65	3111	0.492	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	98.000%		
4) 1,2-Dichloroethane-d4	4.969	65	1525	0.492	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
7) 1,2-Dichloropropane-d6	5.995	67	1841	0.523	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	104.000%		
8) Toluene-d8	7.253	98	3590	0.505	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	102.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1276	0.516	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	104.000%		
Target Compounds						
						Qvalue
3) Vinyl chloride	1.287	62	5937	0.495	ug/L	97
6) Trichloroethene	5.852	95	3004	0.501	ug/L	98
9) 1,2-Dibromoethane	8.305	107	1893	0.504	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	1604	0.512	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.356	75	333	0.518	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
Data File : VV038801.D
Acq On : 16 Jun 2025 12:07
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV247

Quant Time: Jun 16 12:36:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 12:04:36 2025
Response via : Initial Calibration

