

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
 Data File : VV038797.D
 Acq On : 16 Jun 2025 10:15
 Operator : SY/MD
 Sample : VSTD0.143
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1243

Quant Time: Jun 16 11:37:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 11:35:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6431	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	5983	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2671	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.287	65	606	0.101	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.993	65	286	0.093	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.019	67	322	0.087	ug/L	0.02
8) Toluene-d8	7.269	98	621	0.090	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	217	0.069	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.294	62	1116	0.103	ug/L	89
6) Trichloroethene	5.852	95	531	0.098	ug/L	93
9) 1,2-Dibromoethane	8.305	107	357	0.092	ug/L #	90
12) 1,2,3-Trichloropropane	10.214	75	285	0.087	ug/L	95
13) 1,2-Dibromo-3-chloropr...	12.374	75	65	0.089	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
Data File : VV038797.D
Acq On : 16 Jun 2025 10:15
Operator : SY/MD
Sample : VSTD0.143
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1243

Quant Time: Jun 16 11:37:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 11:35:49 2025
Response via : Initial Calibration

