

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
 Data File : VV038799.D
 Acq On : 16 Jun 2025 11:14
 Operator : SY/MD
 Sample : VSTD00145
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001245

Quant Time: Jun 16 11:45:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 16 11:35:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6493	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6105	0.500	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	2656	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	6425	1.058	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.969	65	3317	1.074	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.995	67	3830	1.012	ug/L	0.00
8) Toluene-d8	7.253	98	7700	1.093	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2667	0.834	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.294	62	11750	1.074	ug/L	99
6) Trichloroethene	5.852	95	6326	1.147	ug/L	99
9) 1,2-Dibromoethane	8.290	107	3905	0.991	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	3193	0.978	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	616	0.849	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061625\
Data File : VV038799.D
Acq On : 16 Jun 2025 11:14
Operator : SY/MD
Sample : VSTD00145
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001245

Quant Time: Jun 16 11:45:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 16 11:35:49 2025
Response via : Initial Calibration

