

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM061625.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Jun 16 14:53:20 2025
 Response Via : Initial Calibration

Calibration Files

0.5 =VV038798.D 0.05=VV038796.D 0.1 =VV038797.D 1.0 =VV038799.D 2 =VV038800.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.468	0.474	0.471	0.495	0.467	0.475	2.43
3) T	Vinyl chloride	0.877	0.990	0.868	0.905	0.861	0.900	5.87
4) S	1,2-Dichloroetha...	0.227	0.211	0.222	0.255	0.250	0.233	8.07
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.477	0.483	0.444	0.518	0.492	0.483	5.57
7) S	1,2-Dichloroprop...	0.298	0.243	0.269	0.314	0.292	0.283	9.80
8) S	Toluene-d8	0.583	0.525	0.519	0.631	0.601	0.572	8.51
9) T	1,2-Dibromoethane	0.308	0.282	0.298	0.320	0.303	0.302	4.55
10) S	1,1,2,2-Tetrachl...	0.203	0.190	0.181	0.218	0.202	0.199	7.11
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.557	0.588	0.534	0.601	0.576	0.571	4.64
13) T	1,2-Dibromo-3-ch...	0.109	0.126	0.122	0.116	0.114	0.117	5.63

(#) = Out of Range