

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026379.D
 Acq On : 17 Jun 2022 14:13
 Operator : SY/MD
 Sample : N3318-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJ65

Quant Time: Jun 18 00:04:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 00:00:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.628	114	187439	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	181682	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74921	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	53407	3.280	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	65.600%	
7) Chloroethane-d5	1.587	69	48776	3.810	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	76.200%	
11) 1,1-Dichloroethene-d2	2.127	63	95186	3.405	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.000%	
20) 2-Butanone-d5	3.931	46	84157	52.976	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.960%	
24) Chloroform-d	4.371	84	107708	4.514	ug/L	0.01
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
26) 1,2-Dichloroethane-d4	5.050	65	45635	4.520	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
32) Benzene-d6	5.066	84	202124	4.161	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
36) 1,2-Dichloropropane-d6	6.082	67	65523	4.622	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.400%	
41) Toluene-d8	7.323	98	162701	3.674	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.400%	
43) trans-1,3-Dichloroprop...	7.635	79	13133	2.857	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	57.200%	
46) 2-Hexanone-d5	8.095	63	60710	43.720	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.440%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	41426	4.744	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	55472	4.807	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.200%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.137	96	31010	2.619	ug/L #	76
17) Methyl tert-butyl Ether	2.796	73	16071	0.789	ug/L #	91
19) 1,1-Dichloroethane	3.214	63	6758	0.309	ug/L	88
29) 1,1,1-Trichloroethane	4.628	97	28012	1.395	ug/L	98
34) Trichloroethene	5.924	95	231174	18.179	ug/L	99
47) Tetrachloroethene	7.982	164	6587	0.699	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026379.D
Acq On : 17 Jun 2022 14:13
Operator : SY/MD
Sample : N3318-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJ65

Quant Time: Jun 18 00:04:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 00:00:00 2022
Response via : Initial Calibration

