

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026395.D
 Acq On : 17 Jun 2022 21:16
 Operator : SY/MD
 Sample : N3275-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BD9

Quant Time: Jun 18 01:18:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 01:02:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	177822	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	171462	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70962	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	46269	2.996	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.000%
7) Chloroethane-d5	1.584	69	44971	3.703	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.000%
11) 1,1-Dichloroethene-d2	2.124	63	68699	2.590	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.800%#
20) 2-Butanone-d5	3.924	46	80417	53.360	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.720%
24) Chloroform-d	4.362	84	103667	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.047	65	43403	4.532	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	5.063	84	181949	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.400%
36) 1,2-Dichloropropane-d6	6.079	67	60960	4.556	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%
41) Toluene-d8	7.320	98	141404	3.383	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.600%#
43) trans-1,3-Dichloroprop...	7.632	79	11066	2.551	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	51.000%#
46) 2-Hexanone-d5	8.095	63	59213	45.184	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.360%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42352	5.139	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	52367	4.791	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.800%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.770	101	13240	0.625	ug/L	97
13) Acetone	2.211	43	51740	44.300	ug/L	94
34) Trichloroethene	5.924	95	35019	2.918	ug/L	98
47) Tetrachloroethene	7.979	164	6327	0.712	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026395.D
Acq On : 17 Jun 2022 21:16
Operator : SY/MD
Sample : N3275-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BD9

Quant Time: Jun 18 01:18:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 01:02:22 2022
Response via : Initial Calibration

