

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026396.D
 Acq On : 17 Jun 2022 21:40
 Operator : SY/MD
 Sample : N3278-03 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COBD0

Quant Time: Jun 18 01:18:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 01:02:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	178322	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	173366	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69697	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	48578	3.136	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.800%
7) Chloroethane-d5	1.584	69	46878	3.849	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	2.121	63	71571	2.691	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.800%#
20) 2-Butanone-d5	3.921	46	86586	57.292	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.580%
24) Chloroform-d	4.365	84	110827	4.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
26) 1,2-Dichloroethane-d4	5.047	65	44742	4.658	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
32) Benzene-d6	5.063	84	190137	4.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
36) 1,2-Dichloropropane-d6	6.075	67	63492	4.694	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.320	98	146486	3.466	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.400%#
43) trans-1,3-Dichloroprop...	7.632	79	11835	2.698	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	54.000%#
46) 2-Hexanone-d5	8.091	63	61625	46.508	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.020%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43366	5.204	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	53326	4.967	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	20974	0.987	ug/L	98
16) Methylene chloride	2.523	84	12912	0.898	ug/L	96
34) Trichloroethene	5.934	95	4491	0.370	ug/L	93
47) Tetrachloroethene	7.979	164	83089	9.244	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026396.D
Acq On : 17 Jun 2022 21:40
Operator : SY/MD
Sample : N3278-03 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BD0

Quant Time: Jun 18 01:18:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 01:02:22 2022
Response via : Initial Calibration

