

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026399.D
 Acq On : 17 Jun 2022 22:52
 Operator : SY/MD
 Sample : N3278-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BE8

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/18/2022
 Supervised By :Mahesh Dadoda 06/20/2022

Quant Time: Jun 18 01:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 01:02:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	183482	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	176955	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70796	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	47977	3.010	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	60.200%	
7) Chloroethane-d5	1.584	69	46527	3.713	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	74.200%	
11) 1,1-Dichloroethene-d2	2.121	63	69952	2.556	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	51.200%#	
20) 2-Butanone-d5	3.921	46	94287	60.633	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	121.260%	
24) Chloroform-d	4.365	84	108632	4.651	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
26) 1,2-Dichloroethane-d4	5.047	65	45954	4.650	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
32) Benzene-d6	5.059	84	190810	4.033	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.600%	
36) 1,2-Dichloropropane-d6	6.075	67	61931	4.485	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.800%	
41) Toluene-d8	7.320	98	148051	3.432	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	68.600%#	
43) trans-1,3-Dichloroprop...	7.632	79	11847	2.646	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	53.000%#	
46) 2-Hexanone-d5	8.091	63	59424	43.937	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.880%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43074	5.064	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	54388	4.987	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.800%	
Target Compounds					Qvalue	
13) Acetone	2.214	43	85489	70.938	ug/L	86
21) 2-Butanone	4.031	43	8140m	5.220	ug/L	
34) Trichloroethene	5.924	95	13491	1.089	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026399.D
Acq On : 17 Jun 2022 22:52
Operator : SY/MD
Sample : N3278-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE8

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/18/2022
Supervised By :Mahesh Dadoda 06/20/2022

Quant Time: Jun 18 01:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 01:02:22 2022
Response via : Initial Calibration

