

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026404.D
 Acq On : 18 Jun 2022 00:52
 Operator : SY/MD
 Sample : N3358-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW4R9

Quant Time: Jun 18 01:20:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 01:02:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	175118	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	172058	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67797	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	48753	3.205	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	64.200%	
7) Chloroethane-d5	1.584	69	47112	3.939	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	78.800%	
11) 1,1-Dichloroethene-d2	2.124	63	71509	2.738	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.800%#	
20) 2-Butanone-d5	3.931	46	73775	49.708	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	99.420%	
24) Chloroform-d	4.365	84	111923	5.020	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.400%	
26) 1,2-Dichloroethane-d4	5.047	65	45135	4.785	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
32) Benzene-d6	5.063	84	192910	4.194	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.800%	
36) 1,2-Dichloropropane-d6	6.075	67	64562	4.809	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.200%	
41) Toluene-d8	7.320	98	146333	3.489	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.800%#	
43) trans-1,3-Dichloroprop...	7.632	79	12322	2.830	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	56.600%	
46) 2-Hexanone-d5	8.091	63	63024	47.925	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.860%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44539	5.385	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	52199	4.998	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.310	76	44909	1.646	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026404.D
Acq On : 18 Jun 2022 00:52
Operator : SY/MD
Sample : N3358-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW4R9

Quant Time: Jun 18 01:20:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 01:02:22 2022
Response via : Initial Calibration

