

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061722\
 Data File : VW026402.D
 Acq On : 18 Jun 2022 00:04
 Operator : SY/MD
 Sample : N3358-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW4R6

Quant Time: Jun 18 01:19:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 18 01:02:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	180127	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	173860	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69194	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	47403	3.030	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	60.600%	
7) Chloroethane-d5	1.584	69	46824	3.806	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	76.200%	
11) 1,1-Dichloroethene-d2	2.124	63	69514	2.587	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	51.800%#	
20) 2-Butanone-d5	3.924	46	82981	54.356	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.720%	
24) Chloroform-d	4.365	84	111248	4.851	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
26) 1,2-Dichloroethane-d4	5.047	65	46133	4.755	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.200%	
32) Benzene-d6	5.063	84	190643	4.101	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.000%	
36) 1,2-Dichloropropane-d6	6.079	67	63487	4.680	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.600%	
41) Toluene-d8	7.320	98	147797	3.487	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.800%#	
43) trans-1,3-Dichloroprop...	7.628	79	12738	2.896	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	58.000%	
46) 2-Hexanone-d5	8.091	63	60749	45.717	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.440%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44365	5.309	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	51968	4.876	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061722\
Data File : VV026402.D
Acq On : 18 Jun 2022 00:04
Operator : SY/MD
Sample : N3358-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW4R6

Quant Time: Jun 18 01:19:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 18 01:02:22 2022
Response via : Initial Calibration

