

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061724\
 Data File : VW036097.D
 Acq On : 17 Jun 2024 13:36
 Operator : SY/MD
 Sample : P2856-09
 Misc : 3.65g/10.0mL/MSVOA_V/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0SH1

Quant Time: Jun 18 02:11:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 18 02:10:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	582489	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	596289	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	304696	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	169193	15.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.200%
7) Chloroethane-d5	1.526	69	139452	16.331	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.320%
11) 1,1-Dichloroethene-d2	2.050	65	75623	15.862	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.440%
21) 2-Butanone-d5	3.831	46	64579	26.283	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	52.560%
24) Chloroform-d	4.243	84	297187	16.048	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	64.200%
26) 1,2-Dichloroethane-d4	4.937	65	172427	17.759	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	71.040%
32) Benzene-d6	4.953	84	569738	16.074	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.280%
36) 1,2-Dichloropropane-d6	5.986	67	188383	16.877	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	67.520%#
41) Toluene-d8	7.240	98	476631	15.240	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	60.960%
43) trans-1,3-Dichloroprop...	7.551	79	75886	16.581	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.320%
47) 2-Hexanone-d5	8.031	63	49106	32.471	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.940%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	150335	18.024	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.080%
66) 1,2-Dichlorobenzene-d4	11.558	152	198121	17.731	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	70.920%#
Target Compounds						
					Qvalue	
16) Methylene chloride	2.439	84	154358	11.711	ug/L	97
20) cis-1,2-Dichloroethene	3.818	96	5372	0.540	ug/L	94
46) Tetrachloroethene	7.908	164	9472	1.084	ug/L	98
53) m,p-Xylene	9.079	106	8771	0.547	ug/L	92
54) o-Xylene	9.481	106	11693	0.774	ug/L	99
63) 1,2,4-Trimethylbenzene	10.854	105	11832	0.387	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061724\
Data File : VV036097.D
Acq On : 17 Jun 2024 13:36
Operator : SY/MD
Sample : P2856-09
Misc : 3.65g/10.0mL/MSVOA_V/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0SH1

Quant Time: Jun 18 02:11:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 18 02:10:19 2024
Response via : Initial Calibration

