

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061825\
 Data File : VV038815.D
 Acq On : 18 Jun 2025 12:47
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK224

Quant Time: Jun 18 13:34:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 18 12:17:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6408	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6012	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2543	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.287	65	3078	0.506	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	4.969	65	1526	0.511	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.000%	
7) 1,2-Dichloropropane-d6	5.995	67	1847	0.543	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.253	98	3567	0.519	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1281	0.536	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061825\
Data File : VV038815.D
Acq On : 18 Jun 2025 12:47
Operator : SY/MD
Sample : VIBLK
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK224

Quant Time: Jun 18 13:34:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 18 12:17:29 2025
Response via : Initial Calibration

