

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061825\
 Data File : VV038813.D
 Acq On : 18 Jun 2025 11:49
 Operator : SY/MD
 Sample : Q2275-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 OW-08B-72.5-060925-SIM

Quant Time: Jun 18 12:19:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 18 12:17:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6485	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	5976	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2595	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.287	65	3032	0.492	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.000%
4) 1,2-Dichloroethane-d4	4.970	65	1506	0.498	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	5.996	67	1804	0.533	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.254	98	3430	0.502	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1208	0.508	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061825\
Data File : VV038813.D
Acq On : 18 Jun 2025 11:49
Operator : SY/MD
Sample : Q2275-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
OW-08B-72.5-060925-SIM

Quant Time: Jun 18 12:19:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM061625.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 18 12:17:29 2025
Response via : Initial Calibration

