

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061919\
 Data File : VV011572.D
 Acq On : 18 Jun 2019 13:49
 Operator : SY/MD
 Sample : VV0619WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK35

Quant Time: Jun 19 02:16:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 19 02:08:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	181308	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	169610	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68024	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	57094	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	45826	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.60%
11) 1,1-Dichloroethene-d2	2.12	63	89559	3.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.80%
20) 2-Butanone-d5	3.96	46	171675	44.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.14%
24) Chloroform-d	4.40	84	117972	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	5.08	65	68534	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
32) Benzene-d6	5.09	84	233565	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
36) 1,2-Dichloropropane-d6	6.12	67	76007	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.36	98	203499	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.66	79	22248	4.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.20%
46) 2-Hexanone-d5	8.14	63	128171	41.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.10%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	48811	4.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	67624	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061919\
Data File : VV011572.D
Acq On : 18 Jun 2019 13:49
Operator : SY/MD
Sample : VV0619WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK35

Quant Time: Jun 19 02:16:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 19 02:08:53 2019
Response via : Initial Calibration

