

Data File : VV011584.D
Acq On : 18 Jun 2019 19:28
Operator : SY/MD
Sample : K3289-16DL 4X
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCM1DL

Quant Time: Jun 24 02:00:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 19 02:08:53 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	163676	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	147883	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56163	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55041	4.46	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.20%
7) Chloroethane-d5	1.57	69	37421	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.12	63	91949	3.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.20%
20) 2-Butanone-d5	3.95	46	200809	57.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.50%
24) Chloroform-d	4.40	84	123466	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.08	65	74024	5.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.80%
32) Benzene-d6	5.09	84	238396	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
36) 1,2-Dichloropropane-d6	6.12	67	78138	5.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.60%
41) Toluene-d8	7.36	98	202887	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
43) trans-1,3-Dichloropropene-	7.66	79	23566	5.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.20%
46) 2-Hexanone-d5	8.14	63	149056	55.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	53416	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	64282	5.74	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.80%

Target Compounds

					Qvalue
5) Vinyl chloride	1.32	62	18303	1.172 ug/L	# 78
19) 1,1-Dichloroethane	3.22	63	152540	6.406 ug/L	99
22) cis-1,2-Dichloroethene	3.96	96	80683	6.187 ug/L	96
29) 1,1,1-Trichloroethane	4.65	97	30556	1.661 ug/L	98
34) Trichloroethene	5.96	95	65433	5.312 ug/L	99
47) Tetrachloroethene	8.02	164	10408	1.185 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011584.D
Acq On : 18 Jun 2019 19:28
Operator : SY/MD
Sample : K3289-16DL 4X
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCM1DL

Quant Time: Jun 24 02:00:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 19 02:08:53 2019
Response via : Initial Calibration

