

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061919\
 Data File : VV011606.D
 Acq On : 19 Jun 2019 05:55
 Operator : SY/MD
 Sample : K3386-20
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMB4

Quant Time: Jun 19 09:14:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 19 07:50:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	155888	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	139520	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50545	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	34555	2.94	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.80%
7) Chloroethane-d5	1.58	69	31395	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.20%
11) 1,1-Dichloroethene-d2	2.12	63	56814	2.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.60%#
20) 2-Butanone-d5	3.96	46	184459	55.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.40%
24) Chloroform-d	4.40	84	96365	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
26) 1,2-Dichloroethane-d4	5.08	65	65244	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	5.09	84	186473	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
36) 1,2-Dichloropropane-d6	6.12	67	64757	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.36	98	143155	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.20%
43) trans-1,3-Dichloropropene-	7.67	79	15024	3.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.40%
46) 2-Hexanone-d5	8.14	63	135061	53.23	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.46%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	49783	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	49642	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%

Target Compounds					Ovalue
13) Acetone	2.22	43	6443	2.887 ug/L	78
34) Trichloroethene	5.96	95	2387	0.205 ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061919\
Data File : VV011606.D
Acq On : 19 Jun 2019 05:55
Operator : SY/MD
Sample : K3386-20
Misc : 25ML/MSVOA V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMB4

Quant Time: Jun 19 09:14:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 19 07:50:26 2019
Response via : Initial Calibration

