

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061919\
 Data File : VV011605.D
 Acq On : 19 Jun 2019 05:29
 Operator : SY/MD
 Sample : K3386-19
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMB3

Quant Time: Jun 19 09:10:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 19 07:50:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	153601	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	138762	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	51342	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	35700	3.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.60%
7) Chloroethane-d5	1.58	69	31837	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.60%
11) 1,1-Dichloroethene-d2	2.12	63	58226	2.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.80%#
20) 2-Butanone-d5	3.96	46	187849	57.56	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.12%
24) Chloroform-d	4.40	84	97825	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.08	65	64668	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	5.09	84	190171	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.12	67	66369	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.36	98	141647	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
43) trans-1,3-Dichloropropene-	7.67	79	15931	3.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.80%
46) 2-Hexanone-d5	8.14	63	138515	54.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.78%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	49391	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	50180	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061919\
Data File : VV011605.D
Acq On : 19 Jun 2019 05:29
Operator : SY/MD
Sample : K3386-19
Misc : 25ML/MSVOA V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMB3

Quant Time: Jun 19 09:10:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 19 07:50:26 2019
Response via : Initial Calibration

