

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061920\
 Data File : VV016992.D
 Acq On : 19 Jun 2020 10:23
 Operator : SY/MD
 Sample : VV0619MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK90

Quant Time: Jun 20 03:17:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 20 02:04:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	368275	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	348515	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	168120	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	115049	44.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.54%
7) Chloroethane-d5	1.58	69	79309	50.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.72%
11) 1,1-Dichloroethene-d2	2.12	63	163708	37.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.04%
21) 2-Butanone-d5	3.90	46	162071	84.27	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.27%
24) Chloroform-d	4.37	84	236584	44.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.54%
26) 1,2-Dichloroethane-d4	5.05	65	169247	47.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.18%
32) Benzene-d6	5.07	84	461731	45.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.92%
36) 1,2-Dichloropropane-d6	6.09	67	140740	44.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.88%
41) Toluene-d8	7.33	98	424938	46.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.18%
43) trans-1,3-Dichloropropene-	7.64	79	72826	44.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.30%
47) 2-Hexanone-d5	8.11	63	91264	66.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	66.85%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	162958	39.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	172714	48.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061920\
Data File : VV016992.D
Acq On : 19 Jun 2020 10:23
Operator : SY/MD
Sample : VV0619MBL01
Misc : 5.00µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK90

Quant Time: Jun 20 03:17:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 20 02:04:09 2020
Response via : Initial Calibration

