

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061920\
 Data File : VV017002.D
 Acq On : 19 Jun 2020 15:50
 Operator : SY/MD
 Sample : L3074-03
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9M05

Quant Time: Jun 20 04:30:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 20 02:04:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	320665	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	297788	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	141681	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	106244	46.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.90%
7) Chloroethane-d5	1.58	69	68487	50.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.88%
11) 1,1-Dichloroethene-d2	2.12	63	154584	40.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.38%
21) 2-Butanone-d5	3.91	46	162632	97.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.11%
24) Chloroform-d	4.38	84	224825	48.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.74%
26) 1,2-Dichloroethane-d4	5.06	65	162169	51.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.64%
32) Benzene-d6	5.08	84	439820	51.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.48%
36) 1,2-Dichloropropane-d6	6.09	67	133524	49.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.80%
41) Toluene-d8	7.34	98	400139	50.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.60%
43) trans-1,3-Dichloropropene-	7.64	79	67428	48.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.76%
47) 2-Hexanone-d5	8.11	63	97169	83.29	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.29%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	166934	47.13	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	163160	53.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.92%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	2441	1.068 ug/L	# 21
13) Acetone	2.19	43	3980	3.607 ug/L	91
17) trans-1,2-Dichloroethene	2.78	96	1538	0.709 ug/L	95
20) cis-1,2-Dichloroethene	3.94	96	68723	29.049 ug/L	98
34) Trichloroethene	5.94	95	33768	14.290 ug/L	97
46) Tetrachloroethene	8.00	164	189372	101.898 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061920\
Data File : VV017002.D
Acq On : 19 Jun 2020 15:50
Operator : SY/MD
Sample : L3074-03
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9M05

Quant Time: Jun 20 04:30:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 20 02:04:09 2020
Response via : Initial Calibration

