

Data File : VV016995.D
Acq On : 19 Jun 2020 12:03
Operator : SY/MD
Sample : L3039-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 20 04:02:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 20 02:04:09 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	322675	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	304439	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	150659	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	119211	52.36	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	104.72%		
7) Chloroethane-d5	1.58	69	76316	55.86	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	111.72%		
11) 1,1-Dichloroethene-d2	2.12	63	168848	44.16	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.32%		
21) 2-Butanone-d5	3.90	46	170146	100.97	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.97%		
24) Chloroform-d	4.37	84	244304	52.77	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	105.54%		
26) 1,2-Dichloroethane-d4	5.05	65	172669	54.83	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	109.66%		
32) Benzene-d6	5.07	84	478291	54.50	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	109.00%		
36) 1,2-Dichloropropane-d6	6.09	67	145432	53.16	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	106.32%		
41) Toluene-d8	7.33	98	437636	54.34	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.68%		
43) trans-1,3-Dichloropropene-	7.64	79	73112	51.32	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	102.64%		
47) 2-Hexanone-d5	8.11	63	93669	78.54	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	78.54%		
57) 1,1,2,2-Tetrachloroethane-	10.23	84	166257	45.91	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.82%		
64) 1,2-Dichlorobenzene-d4	11.65	152	179023	55.68	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	111.36%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016995.D
Acq On : 19 Jun 2020 12:03
Operator : SY/MD
Sample : L3039-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 20 04:02:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 20 02:04:09 2020
Response via : Initial Calibration

