

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062019\
 Data File : VV011623.D
 Acq On : 19 Jun 2019 17:41
 Operator : SY/MD
 Sample : K3386-09DL 5X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM88DL

Quant Time: Jun 20 04:02:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 20 02:04:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	154414	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	143746	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	58171	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	58615	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.57	69	34971	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.60%
11) 1,1-Dichloroethene-d2	2.11	63	98289	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.60%
20) 2-Butanone-d5	3.96	46	161121	49.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.22%
24) Chloroform-d	4.40	84	109147	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.40%
26) 1,2-Dichloroethane-d4	5.08	65	66190	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
32) Benzene-d6	5.09	84	232176	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.11	67	71606	5.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.20%
41) Toluene-d8	7.36	98	205726	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
43) trans-1,3-Dichloropropene-	7.66	79	21838	4.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.40%
46) 2-Hexanone-d5	8.14	63	115866	44.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	46787	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	65390	5.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.60%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	12527	1.304 ug/L	# 16
19) 1,1-Dichloroethane	3.22	63	2571	0.114 ug/L	98
22) cis-1,2-Dichloroethene	3.95	96	17600	1.431 ug/L	96
29) 1,1,1-Trichloroethane	4.65	97	5240	0.293 ug/L	95
34) Trichloroethene	5.96	95	197803	16.520 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062019\
Data File : VV011623.D
Acq On : 19 Jun 2019 17:41
Operator : SY/MD
Sample : K3386-09DL 5X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM88DL

Quant Time: Jun 20 04:02:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 20 02:04:09 2019
Response via : Initial Calibration

