

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062022\
 Data File : VW026440.D
 Acq On : 20 Jun 2022 12:54
 Operator : SY/MD
 Sample : N3358-14
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW4S7

Quant Time: Jun 21 02:33:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 21 02:31:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	221250	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	203783	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	83499	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	78400	4.080	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.587	69	66977	4.433	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.124	63	113941	3.453	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.000%
20) 2-Butanone-d5	3.937	46	88998	47.462	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.920%
24) Chloroform-d	4.365	84	143523	5.096	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.047	65	56609	4.750	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	5.063	84	295874	5.431	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.600%
36) 1,2-Dichloropropane-d6	6.079	67	84216	5.296	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.320	98	254073	5.114	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
43) trans-1,3-Dichloroprop...	7.632	79	16593	3.218	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.400%
46) 2-Hexanone-d5	8.095	63	69515	44.632	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.260%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	48042	4.905	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	77026	5.989	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062022\
Data File : VV026440.D
Acq On : 20 Jun 2022 12:54
Operator : SY/MD
Sample : N3358-14
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW4S7

Quant Time: Jun 21 02:33:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 21 02:31:12 2022
Response via : Initial Calibration

