

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062022\
 Data File : VW026448.D
 Acq On : 20 Jun 2022 16:12
 Operator : SY/MD
 Sample : N3256-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK3

Quant Time: Jun 21 02:35:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 21 02:31:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	194288	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	181400	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76438	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	71553	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.587	69	61019	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.00%
11) 1,1-Dichloroethene-d2	2.124	63	103764	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	3.928	46	84191	51.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.26%
24) Chloroform-d	4.362	84	135276	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.40%
26) 1,2-Dichloroethane-d4	5.043	65	51903	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.059	84	267779	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
36) 1,2-Dichloropropane-d6	6.075	67	78328	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.317	98	224841	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
43) trans-1,3-Dichloroprop...	7.628	79	16714	3.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.80%
46) 2-Hexanone-d5	8.091	63	64689	46.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.32%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	45168	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
66) 1,2-Dichlorobenzene-d4	11.622	152	69928	5.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.80%
Target Compounds						
16) Methylene chloride	2.526	84	7019	0.45	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062022\
Data File : VV026448.D
Acq On : 20 Jun 2022 16:12
Operator : SY/MD
Sample : N3256-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AK3

Quant Time: Jun 21 02:35:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 21 02:31:12 2022
Response via : Initial Calibration

