

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062022\
 Data File : VV026451.D
 Acq On : 20 Jun 2022 17:25
 Operator : SY/MD
 Sample : VV0620WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK381

Quant Time: Jun 21 05:26:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 21 05:26:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	184475	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	175856	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74529	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.326	65	61493	3.838	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.587	69	55587	4.412	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.200%
11) 1,1-Dichloroethene-d2	2.124	63	92201	3.351	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.000%
20) 2-Butanone-d5	3.924	46	91099	58.268	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.540%
24) Chloroform-d	4.365	84	120253	5.120	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
26) 1,2-Dichloroethane-d4	5.043	65	49886	5.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.059	84	242513	5.158	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
36) 1,2-Dichloropropane-d6	6.075	67	72560	5.288	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.800%
41) Toluene-d8	7.320	98	200320	4.673	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
43) trans-1,3-Dichloroprop...	7.628	79	15520	3.488	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.800%
46) 2-Hexanone-d5	8.091	63	69645	51.816	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.640%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46106	5.454	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	65566	5.711	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062022\
Data File : VV026451.D
Acq On : 20 Jun 2022 17:25
Operator : SY/MD
Sample : VV0620WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK381

Quant Time: Jun 21 05:26:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 21 05:26:17 2022
Response via : Initial Calibration

