

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062024\
 Data File : VV036160.D
 Acq On : 20 Jun 2024 13:37
 Operator : SY/MD
 Sample : P2909-20
 Misc : 5.32g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Jun 21 01:48:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 20 05:02:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	539644	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	462796	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	128586	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	143292	14.443	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.760%
7) Chloroethane-d5	1.526	69	127402	16.105	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.400%
11) 1,1-Dichloroethene-d2	2.050	65	68015	15.399	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.600%
21) 2-Butanone-d5	3.838	46	44798	19.680	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	39.360%
24) Chloroform-d	4.243	84	275551	16.061	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	64.240%
26) 1,2-Dichloroethane-d4	4.937	65	142589	15.852	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	63.400%#
32) Benzene-d6	4.953	84	503720	18.311	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.240%
36) 1,2-Dichloropropane-d6	5.986	67	167923	19.384	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.520%
41) Toluene-d8	7.239	98	388213	15.994	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	63.960%
43) trans-1,3-Dichloroprop...	7.555	79	52005	14.641	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	58.560%
47) 2-Hexanone-d5	8.030	63	30882	26.311	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	52.620%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	98264	15.179	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	60.720%
66) 1,2-Dichlorobenzene-d4	11.558	152	79712	16.905	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	67.600%#
Target Compounds						
					Qvalue	
13) Acetone	2.143	43	6270	1.768	ug/L #	1
16) Methylene chloride	2.442	84	31321	2.565	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062024\
Data File : VV036160.D
Acq On : 20 Jun 2024 13:37
Operator : SY/MD
Sample : P2909-20
Misc : 5.32g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Jun 21 01:48:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 20 05:02:48 2024
Response via : Initial Calibration

