

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
 Data File : VV011642.D
 Acq On : 20 Jun 2019 15:09
 Operator : SY/MD
 Sample : K3386-17
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMAO

Quant Time: Jun 21 02:18:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 01:58:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	169800	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	151933	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	60305	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	59777	5.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.00%
7) Chloroethane-d5	1.58	69	40075	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.12	63	103486	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.40%
20) 2-Butanone-d5	3.95	46	160052	53.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.82%
24) Chloroform-d	4.40	84	111806	5.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.80%
26) 1,2-Dichloroethane-d4	5.08	65	65823	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.09	84	228366	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
36) 1,2-Dichloropropane-d6	6.12	67	74078	5.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.60%
41) Toluene-d8	7.36	98	193258	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
43) trans-1,3-Dichloropropene-	7.66	79	21674	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
46) 2-Hexanone-d5	8.14	63	120452	56.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47806	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	62352	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	2.13	96	31940	3.039	ug/L #	83
13) Acetone	2.21	43	12725	5.934	ug/L	96
22) cis-1,2-Dichloroethene	3.96	96	32236	2.374	ug/L #	94
29) 1,1,1-Trichloroethane	4.65	97	15028	0.768	ug/L	99
34) Trichloroethene	5.96	95	49154	3.680	ug/L	97
42) Toluene	7.43	91	12589	0.239	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
Data File : VV011642.D
Acq On : 20 Jun 2019 15:09
Operator : SY/MD
Sample : K3386-17
Misc : 25ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMAO

Quant Time: Jun 21 02:18:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 01:58:40 2019
Response via : Initial Calibration

