

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
 Data File : VV011644.D
 Acq On : 20 Jun 2019 15:56
 Operator : SY/MD
 Sample : VIBLK25
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK25

Quant Time: Jun 21 02:26:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 01:58:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	148212	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	136023	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56790	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	58234	5.92	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.40%
7) Chloroethane-d5	1.57	69	42075	6.08	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.60%
11) 1,1-Dichloroethene-d2	2.12	63	86714	4.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.80%
20) 2-Butanone-d5	3.95	46	145107	56.00	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.00%
24) Chloroform-d	4.40	84	109315	6.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.80%
26) 1,2-Dichloroethane-d4	5.08	65	62232	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.40%
32) Benzene-d6	5.10	84	224455	5.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.40%
36) 1,2-Dichloropropane-d6	6.12	67	69748	5.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.40%
41) Toluene-d8	7.36	98	192254	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
43) trans-1,3-Dichloropropene-	7.67	79	19416	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
46) 2-Hexanone-d5	8.14	63	113126	59.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.08%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	44195	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	63000	5.88	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062119\
Data File : VV011644.D
Acq On : 20 Jun 2019 15:56
Operator : SY/MD
Sample : VIBLK25
Misc : 25ML/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK25

Quant Time: Jun 21 02:26:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 01:58:40 2019
Response via : Initial Calibration

