

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
 Data File : VV011663.D
 Acq On : 21 Jun 2019 00:15
 Operator : SY/MD
 Sample : K3415-09
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GALP6

Manual Integrations
 APPROVED

MMDadoda
 6/21/2019 9:42:18 AM

Quant Time: Jun 21 04:20:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 02:33:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	154371	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	142138	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56815	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52543	5.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.60%
7) Chloroethane-d5	1.56	69	36307	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.12	63	80389	3.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.20%
20) 2-Butanone-d5	3.95	46	160088	59.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.62%
24) Chloroform-d	4.40	84	89525	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.08	65	62426	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.20%
32) Benzene-d6	5.09	84	209146	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.12	67	65740	5.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.60%
41) Toluene-d8	7.36	98	178912	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.66	79	19654	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.14	63	117592	58.73	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.46%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47030	5.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	59516	5.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.00%

Target Compounds

					Ovalue
13) Acetone	2.20	43	27211	13.957 ug/L	98
21) 2-Butanone	4.05	43	5341m	1.700 ug/L	
25) Chloroform	4.42	83	32014	1.306 ug/L	96
31) Carbon tetrachloride	4.87	117	6676329	431.555 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
Data File : VV011663.D
Acq On : 21 Jun 2019 00:15
Operator : SY/MD
Sample : K3415-09
Misc : 25ML/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GALP6

Manual Integrations
APPROVED

MMDadoda
6/21/2019 9:42:18 AM

Quant Time: Jun 21 04:20:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 02:33:37 2019
Response via : Initial Calibration

