

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
 Data File : VV011671.D
 Acq On : 21 Jun 2019 03:38
 Operator : SY/MD
 Sample : K3381-17 2000X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCM7

Quant Time: Jun 21 04:57:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 04:33:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	159099	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	152844	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	60002	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	45755	4.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	1.58	69	30146	4.06	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.20%
11) 1,1-Dichloroethene-d2	2.12	63	73360	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	3.96	46	152127	54.69	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.38%
24) Chloroform-d	4.40	84	75081	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.00%
26) 1,2-Dichloroethane-d4	5.08	65	56837	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.09	84	190612	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.40%
36) 1,2-Dichloropropane-d6	6.12	67	59719	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.36	98	160591	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
43) trans-1,3-Dichloropropene-	7.66	79	17388	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.14	63	111454	51.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	42238	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	53781	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds

					Ovalue
19) 1,1-Dichloroethane	3.22	63	9720	0.422 ug/L	96
22) cis-1,2-Dichloroethene	3.96	96	11016	0.866 ug/L	97
25) Chloroform	4.42	83	21203	0.839 ug/L	98
29) 1,1,1-Trichloroethane	4.65	97	51811	2.633 ug/L	100
34) Trichloroethene	5.96	95	2932	0.218 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062119\
Data File : VV011671.D
Acq On : 21 Jun 2019 03:38
Operator : SY/MD
Sample : K3381-17 2000X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCM7

Quant Time: Jun 21 04:57:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 04:33:10 2019
Response via : Initial Calibration

