

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062119\
 Data File : VV011640.D
 Acq On : 20 Jun 2019 13:32
 Operator : SY/MD
 Sample : VV0621WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK38

Quant Time: Jun 21 02:11:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 21 01:58:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	158373	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	148993	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	62011	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	57148	5.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.60%
7) Chloroethane-d5	1.58	69	44123	5.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.40%
11) 1,1-Dichloroethene-d2	2.12	63	89499	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.80%
20) 2-Butanone-d5	3.96	46	175290	63.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.62%
24) Chloroform-d	4.40	84	114173	5.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	119.00%
26) 1,2-Dichloroethane-d4	5.08	65	67702	5.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.40%
32) Benzene-d6	5.09	84	237295	5.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.20%
36) 1,2-Dichloropropane-d6	6.11	67	71391	5.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.60%
41) Toluene-d8	7.36	98	204124	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.60%
43) trans-1,3-Dichloropropene-	7.66	79	23776	5.32	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.40%
46) 2-Hexanone-d5	8.14	63	132886	63.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	126.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	51646	5.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	67480	5.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062119\
Data File : VV011640.D
Acq On : 20 Jun 2019 13:32
Operator : SY/MD
Sample : VV0621WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK38

Quant Time: Jun 21 02:11:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 21 01:58:40 2019
Response via : Initial Calibration

