

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062123\
 Data File : VV031601.D
 Acq On : 21 Jun 2023 12:08
 Operator : SY/MD
 Sample : VV0621WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK360

Quant Time: Jun 22 02:14:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 22 02:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	206782	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	202425	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	97004	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	76611	55.669	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	111.340%
7) Chloroethane-d5	1.539	69	65686	54.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.400%
11) 1,1-Dichloroethene-d2	2.063	65	36486	51.442	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.880%
21) 2-Butanone-d5	3.802	46	98279	88.194	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.190%
24) Chloroform-d	4.256	84	156078	45.285	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.560%
26) 1,2-Dichloroethane-d4	4.947	65	107917	50.045	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.080%
32) Benzene-d6	4.966	84	294613	51.706	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.420%
36) 1,2-Dichloropropane-d6	5.992	67	101195	50.109	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.220%
41) Toluene-d8	7.249	98	244536	46.769	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.540%
43) trans-1,3-Dichloroprop...	7.558	79	38851	39.463	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.920%
47) 2-Hexanone-d5	8.030	63	72939	91.476	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.480%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	140234	46.005	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.020%
66) 1,2-Dichlorobenzene-d4	11.567	152	99603	53.867	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062123\
Data File : VV031601.D
Acq On : 21 Jun 2023 12:08
Operator : SY/MD
Sample : VV0621WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK360

Quant Time: Jun 22 02:14:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 22 02:14:12 2023
Response via : Initial Calibration

