

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062123\
 Data File : VW031603.D
 Acq On : 21 Jun 2023 12:55
 Operator : SY/MD
 Sample : 03234-15ME
 Misc : 5.52g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YBTY9ME

Quant Time: Jun 22 02:15:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 22 02:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	249138	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	250003	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.197	152	126404	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	74512	44.938	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	89.880%		
7) Chloroethane-d5	1.532	69	37084	25.633	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	51.260%#		
11) 1,1-Dichloroethene-d2	2.047	65	38227	44.733	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	89.460%		
21) 2-Butanone-d5	3.825	46	82227	61.244	ug/L	0.03
Spiked Amount 100.000	Range 40 - 130		Recovery =	61.240%		
24) Chloroform-d	4.259	84	113436	27.317	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	54.640%#		
26) 1,2-Dichloroethane-d4	4.947	65	105178	40.482	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	80.960%		
32) Benzene-d6	4.963	84	300966	42.769	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	85.540%		
36) 1,2-Dichloropropane-d6	5.995	67	100118	40.141	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	80.280%		
41) Toluene-d8	7.249	98	263098	40.742	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	81.480%		
43) trans-1,3-Dichloroprop...	7.561	79	47481	39.051	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.100%		
47) 2-Hexanone-d5	8.069	63	65242	66.251	ug/L	0.04
Spiked Amount 100.000	Range 45 - 130		Recovery =	66.250%		
56) 1,1,2,2-Tetrachloroeth...	10.169	84	133960	35.583	ug/L	0.01
Spiked Amount 50.000	Range 65 - 120		Recovery =	71.160%		
66) 1,2-Dichlorobenzene-d4	11.570	152	104156	43.228	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.460%		
Target Compounds					Qvalue	
25) Chloroform	4.288	83	51526	11.935	ug/L	98
34) Trichloroethene	5.838	95	31231	14.042	ug/L	90
46) Tetrachloroethene	7.911	164	1609	0.922	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062123\
Data File : VV031603.D
Acq On : 21 Jun 2023 12:55
Operator : SY/MD
Sample : 03234-15ME
Misc : 5.52g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YBTY9ME

Quant Time: Jun 22 02:15:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM061423WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 22 02:14:12 2023
Response via : Initial Calibration

