

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062220\
 Data File : VV017013.D
 Acq On : 22 Jun 2020 11:37
 Operator : SY/MD
 Sample : VV0622WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK92

Quant Time: Jun 23 03:45:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 23 03:22:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	336769	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	318018	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	152117	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	104959	44.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.34%
7) Chloroethane-d5	1.58	69	71081	49.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.70%
11) 1,1-Dichloroethene-d2	2.12	63	156437	39.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.40%
21) 2-Butanone-d5	3.91	46	162218	92.23	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.23%
24) Chloroform-d	4.38	84	233522	48.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.66%
26) 1,2-Dichloroethane-d4	5.06	65	164553	50.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.14%
32) Benzene-d6	5.08	84	452683	49.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.76%
36) 1,2-Dichloropropane-d6	6.09	67	140494	49.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.34%
41) Toluene-d8	7.34	98	409931	48.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.46%
43) trans-1,3-Dichloropropene-	7.64	79	71251	47.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.74%
47) 2-Hexanone-d5	8.11	63	92944	74.60	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.60%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	164468	43.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.96%
64) 1,2-Dichlorobenzene-d4	11.65	152	166821	51.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.78%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062220\
Data File : VV017013.D
Acq On : 22 Jun 2020 11:37
Operator : SY/MD
Sample : VV0622WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK92

Quant Time: Jun 23 03:45:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 23 03:22:30 2020
Response via : Initial Calibration

