




Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062320\
Data File : VV017038.D
Acq On : 22 Jun 2020 23:45
Operator : SY/MD
Sample : L3067-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jun 23 05:00:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 23 04:53:49 2020
Response via : Initial Calibration



TIC: VV017038.D

(12) 1,1-Dichloroethene (T)

2.135min (+0.003) 0.59ug/L m

response 4897

Ion	Exp%	Act%
96.00	100	100
61.05	178.60	226.70
63.00	156.80	614.48#
0.00	0.00	0.00