

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062320\
 Data File : VV017058.D
 Acq On : 23 Jun 2020 16:10
 Operator : SY/MD
 Sample : L3067-19DL2 1000X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXH6DL2

Quant Time: Jun 24 06:59:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 24 05:01:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	172082	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	167603	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	79917	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43575	4.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.40%
7) Chloroethane-d5	1.58	69	35402	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.12	63	66783	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.93	46	127493	48.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.88%
24) Chloroform-d	4.38	84	97767	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.06	65	57684	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.07	84	202141	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	6.09	67	60960	4.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.60%
41) Toluene-d8	7.34	98	183216	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
45) trans-1,3-Dichloropropene-	7.64	79	23466	4.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.60%
48) 2-Hexanone-d5	8.11	63	78810	38.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.30%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	40002	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	72331	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	2777	0.336 ug/L	# 1
19) 1,1-Dichloroethane	3.21	63	39129	1.990 ug/L	99
25) Chloroform	4.40	83	11676	0.544 ug/L	94
29) 1,1,1-Trichloroethane	4.63	97	140991	7.021 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062320\
Data File : VV017058.D
Acq On : 23 Jun 2020 16:10
Operator : SY/MD
Sample : L3067-19DL2 1000X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXH6DL2

Quant Time: Jun 24 06:59:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 24 05:01:54 2020
Response via : Initial Calibration

