

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062320\
 Data File : VV017028.D
 Acq On : 22 Jun 2020 19:45
 Operator : SY/MD
 Sample : L3067-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXG4

Quant Time: Jun 23 05:50:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 23 04:53:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	184741	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	178101	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	88347	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	47045	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.58	69	37029	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.13	63	72606	3.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.80%
20) 2-Butanone-d5	3.93	46	136276	48.22	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.44%
24) Chloroform-d	4.38	84	112513	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.06	65	61973	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.07	84	217299	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.09	67	64529	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.20%
41) Toluene-d8	7.34	98	201459	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
45) trans-1,3-Dichloropropene-	7.64	79	22535	3.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.20%
48) 2-Hexanone-d5	8.11	63	84015	38.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.54%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	43961	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	81292	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062320\
Data File : VV017028.D
Acq On : 22 Jun 2020 19:45
Operator : SY/MD
Sample : L3067-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXG4

Quant Time: Jun 23 05:50:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 23 04:53:49 2020
Response via : Initial Calibration

