

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026526.D
 Acq On : 23 Jun 2022 19:55
 Operator : SY/MD
 Sample : N3400-19
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF4

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/28/2022
 Supervised By :Mahesh Dadoda 06/28/2022

Quant Time: Jun 24 02:05:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 24 00:51:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	229793	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	215482	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	93629	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	86530	4.682	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.600%	
7) Chloroethane-d5	1.577	69	70035	4.717	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.400%	
11) 1,1-Dichloroethene-d2	2.117	63	114314	3.481	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.600%	
20) 2-Butanone-d5	3.928	46	110073	59.910	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	119.820%	
24) Chloroform-d	4.358	84	138561	4.587	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.800%	
26) 1,2-Dichloroethane-d4	5.040	65	58188	5.001	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.000%	
32) Benzene-d6	5.056	84	289666	4.680	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.600%	
36) 1,2-Dichloropropane-d6	6.075	67	83752	4.921	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.400%	
41) Toluene-d8	7.317	98	252308	4.533	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.600%	
43) trans-1,3-Dichloroprop...	7.628	79	16481	3.746	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.000%	
46) 2-Hexanone-d5	8.091	63	87975	52.500	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	105.000%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	54094	4.858	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	80015	5.164	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.200%	
Target Compounds					Qvalue	
13) Acetone	2.223	43	55004	47.190	ug/L	79
21) 2-Butanone	4.018	43	6907m	3.939	ug/L	
30) Cyclohexane	4.683	56	4150	0.198	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026526.D
Acq On : 23 Jun 2022 19:55
Operator : SY/MD
Sample : N3400-19
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF4

Quant Time: Jun 24 02:05:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 24 00:51:15 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/28/2022
Supervised By :Mahesh Dadoda 06/28/2022

