

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062322\
 Data File : VW026535.D
 Acq On : 23 Jun 2022 23:30
 Operator : SY/MD
 Sample : N3474-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA6

Quant Time: Jun 24 02:07:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 24 00:51:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	237136	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	219392	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	95251	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	82656	4.334	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.600%
7) Chloroethane-d5	1.577	69	68421	4.466	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	2.117	63	111180	3.281	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.924	46	107370	56.629	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.260%
24) Chloroform-d	4.358	84	140486	4.506	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.040	65	56740	4.725	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
32) Benzene-d6	5.056	84	291060	4.619	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.075	67	84628	4.884	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.317	98	243019	4.288	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	7.628	79	15821	3.532	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.600%
46) 2-Hexanone-d5	8.091	63	86976	50.979	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.960%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	53752	4.742	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	79536	5.046	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%
Target Compounds						
13) Acetone	2.211	43	8749	7.274	ug/L	60
16) Methylene chloride	2.519	84	9513	0.560	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062322\
Data File : VV026535.D
Acq On : 23 Jun 2022 23:30
Operator : SY/MD
Sample : N3474-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA6

Quant Time: Jun 24 02:07:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR062322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 24 00:51:15 2022
Response via : Initial Calibration

